

# **EC Declaration of Conformity**

In accordance with EN ISO 17050-1:2010



### The product manufacturer and location of Technical Construction File,

Setra Systems, Inc

159 Swanson Road, Boxborough, MA 01719, USA Tel: +1-978-263-1400 Fax: +1-978-264-0292

## and the authorized representative established within the Community,

Hengstler GmbH

Uhlandstr. 49, 78554 Aldingen, Germany Tel: +49-7424-890, Fax +49-7424-89500

### herewith declare that the products as originally delivered:

Product Name and Model:

Setra Power Patrol (model SPP)

Model number:

**SPPxx** 

(x=alphanumerical numbers for different configurations)

# comply with the essential requirements of the following applicable European Directives and carry the CE marking accordingly:

EMC Directive (2004/108/EC)

LVD (Low Voltage) Directive (2006/95/EC)

RoHS Directive (2002/95/EU)

### The following standards were used to demonstrate compliance:

CISPR 22: 2008

USA: FCC 15.109(g): 2014

USA: FCC 15.107:2014

EN61326-1:2006: 2006

EN 61000-4-2

EN 61000-4-3

EN 61000-4-4

EN 61000-4-5

EN 61000-4-6

EN 61000-4-8

EN61000-3-3:2008

EN61000-6-4:2007+A1:2011

GROUP 1, CLASS A EMISSION LIMITS

Radiated Emissions

Conducted Emissions

### INDUSTRIAL IMMUNITY REQUIREMENTS

4 kV contact, 8 kV air

10 V/m 80-1000MHz, 3 V/m 1.4-2GHz, 1 V/m 2-2.7GHz

2 kV power lines, 1 kV signal/control lines

2 kV line-line, 2 line-neutral

 $3V_{rms}$ , 0.15-80MHz, 1 kHz, 80% AM

Rated power frequency magnetic field, 30 A/m

EMC. Limitation of voltage changes, voltage fluctuations and flicker in

public low-voltage supply system

EMC Emissions standard for industrial environments

### Safety Standard

IEC/EN61010-1, 3<sup>rd</sup> Edition UL 61010-1, 3<sup>rd</sup> Edition CAN/CSA-C22.2 No. 61010-1, 3<sup>rd</sup> Edition

IEC 61010-2-030, 1st Edition

#### Description

Safety Requirements for Electrical Equipment for Measurement Control, and Laboratory Use, Part 1: General Requirements

Safety Requirements for Electrical Equipment for Measurement Control, and Laboratory Use, Part 2-030: Particular Requirements for Testing and Measuring Circuits

DoC Number: CE-SPP-A

I hereby declare that the product(s) named above have been designed to comply with the relevant sections of the above referenced standards. The product(s) comply with all applicable Essential Requirements of the Directives. Subject to correct installation, maintenance and use conforming to its (their) intended purpose, to the applicable regulations and standards, to our operation and maintenance manual, the product complies with the provision of the above mentioned European Directives.

May 15, 2014

Eric Beishline

**VP** Engineering